

**Declaration of conformity**

**CE 0123**

Manufacturer Address Ivoclar Vivadent AG  
 Bendererstrasse 2  
 FL – 9494 Schaan  
 Liechtenstein

Product **IPS e.max® Press**

Type of material Ceramic system

Classification Medical Device Class IIa

We herewith declare that the above mentioned products meet the provisions of the following EC Council Directives and its implementation in national law. All supporting documentations are retained under the premises of the manufacturer and the notified body.

Directives Medical Device Directive 93/42/EEC, Annex II. 3

Standards ISO 6872 Dental ceramic  
 ISO 9693 Metal-ceramic dental restorative systems

Notified Body Address TÜV Product Services  
 Ridlerstrasse 65  
 DE – 80339 München  
 Deutschland

Place, Valid from Schaan, 08 May 2007  
 Replaces version of: 03 October 2005

Signature 

Name Position Dr. Volker Rheinberger CTO Dipl. Ing. Patrik Oehri Director R&D Services

Date 07 May 2007 07 May 2007